## Rat Urotensin-2 receptor (UTS2R) ELISA Kit

Catalog No: #EK5921

Package Size: #EK5921-1 48T #EK5921-2 96T



Orders: order@signalwayantibody.com Support: tech@signalwayantibody.com

## Description

Product NameRat Urotensin-2 receptor (UTS2R) ELISA KitBrief DescriptionELISA KitApplicationsELISASpecies ReactivityRat (Rattus norvegicus)Other NamesGPR14; UTR; UTR2; G protein-coupled receptor 14Accession No.P49684UniprotP49684GenelD57305;	
ApplicationsELISASpecies ReactivityRat (Rattus norvegicus)Other NamesGPR14; UTR; UTR2; G protein-coupled receptor 14Accession No.P49684UniprotP49684GeneID57305;	
Species ReactivityRat (Rattus norvegicus)Other NamesGPR14; UTR; UTR2; G protein-coupled receptor 14Accession No.P49684UniprotP49684GeneID57305;	
Other NamesGPR14; UTR; UTR2; G protein-coupled receptor 14Accession No.P49684UniprotP49684GeneID57305;	
Accession No.P49684UniprotP49684GeneID57305;	
UniprotP49684GeneID57305;	
GenelD 57305;	
Storage The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less t	han 5%
within the expiration date under appropriate storage condition.	
The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 a	and 7 days,
and compare O.D.values of the kit kept at 37C with that of at recommended temperature. (referring	ng from China
Biological Products Standard, which was calculated by the Arrhenius equation. For ELISA kit, 4 d	ays storage
at 37C can be considered as 6 months at 2 - 8C, which means 7 days at 37C equaling 12 months	s at 2 - 8C).

Application Details	
Detect Range:0.312-20 ng/mL	
Sensitivity:0.108 ng/mL	
Sample Type:Serum, Plasma,	Other biological fluids
Sample Volume: 1-200 µL	
Assay Time:1-4.5h	
Detection wavelength:450 nm	

## Product Description

Detection Method:SandwichTest principle:This assay employs a two-site sandwich ELISA to quantitate UTS2R in samples. An antibody specific for UTS2R has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyUTS2R present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for UTS2R is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of UTS2R bound in the initial step. The color development is stopped and the intensity of the color is measured.

Note: This product is for in vitro research use only