Canine Tapasin (TAPBP) ELISA Kit

Catalog No: #EK6941

Package Size: #EK6941-1 48T #EK6941-2 96T



Orders: order@signalwayantibody.com Support: tech@signalwayantibody.com

$\overline{}$		4.0
	escri	ntion
u	COUL	UUUII

Product Name	Canine Tapasin (TAPBP) ELISA Kit	
Brief Description	ELISA Kit	
Applications	ELISA	
Species Reactivity	Canine (Canis familiaris; Dog)	
Other Names	DADB-159G18.7; NGS17; TAPA; TPN; TPSN; TAP-associated protein tapasin	
Accession No.	Q5TJE4	
Uniprot	Q5TJE4	
GeneID	481740;	
Storage	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5%	
	within the expiration date under appropriate storage condition.	
	The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days,	
	and compare O.D.values of the kit kept at 37C with that of at recommended temperature. (referring from China	
	Biological Products Standard, which was calculated by the Arrhenius equation. For ELISA kit, 4 days storage	
	at 37C can be considered as 6 months at 2 - 8C, which means 7 days at 37C equaling 12 months at 2 - 8C).	

Application Details

Detect Range:Request Information		
Sensitivity:Request Information		
Sample Type:Serum, Plasma, Other biological fluids		
Sample Volume: 1-200 μL		
Assay Time:1-4.5h		
Detection wavelength:450 nm		

Product Description

Detection Method:SandwichTest principle:This assay employs a two-site sandwich ELISA to quantitate TAPBP in samples. An antibody specific for TAPBP has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyTAPBP present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for TAPBP is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of TAPBP bound in the initial step. The color development is stopped and the intensity of the color is measured.

Note: This product is for in vitro research use only