Pig Transforming growth factor α (TGFα) ELISA Kit

Catalog No: #EK7042

Package Size: #EK7042-1 48T #EK7042-2 96T



Orders: order@signalwayantibody.com Support: tech@signalwayantibody.com

$\overline{}$			
	escri	nti	n
\boldsymbol{u}	COUL	ĮΟU	OI.

Product Name	Pig Transforming growth factor α (TGF α) ELISA Kit	
Brief Description	ELISA Kit	
Applications	ELISA	
Species Reactivity	Pig (Sus scrofa; Porcine)	
Other Names	TFGA; transforming growth factor-alpha	
Accession No.	Q06922	
Uniprot	Q06922	
GeneID	397484;	
Storage	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5%	
	within the expiration date under appropriate storage condition.	
	The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days,	
	and compare O.D.values of the kit kept at 37C with that of at recommended temperature. (referring from China	
	Biological Products Standard, which was calculated by the Arrhenius equation. For ELISA kit, 4 days storage	
	at 37C can be considered as 6 months at 2 - 8C, which means 7 days at 37C equaling 12 months at 2 - 8C).	

Application Details

Detect Range:Request Information		
Sensitivity:Request Information		
Sample Type:Serum, Plasma, Other biological fluids		
Sample Volume: 1-200 μL		
Assay Time:1-4.5h		
Detection wavelength:450 nm		

Product Description

Detection Method:SandwichTest principle:This assay employs a two-site sandwich ELISA to quantitate TGFA in samples. An antibody specific for TGFA has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyTGFA present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for TGFA is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of TGFA bound in the initial step. The color development is stopped and the intensity of the color is measured.

Note: This product is for in vitro research use only