Rat 3-nitrotyrosine (3-NT) ELISA Kit

Catalog No: #EK7578

Package Size: #EK7578-1 48T #EK7578-2 96T



Orders: order@signalwayantibody.com Support: tech@signalwayantibody.com

Description

| Product Name | Rat 3-nitrotyrosine (3-NT) ELISA Kit | | |
|--------------------|--|--|--|
| Brief Description | ELISA Kit | | |
| Applications | ELISA | | |
| Species Reactivity | Rat (Rattus norvegicus) | | |
| Storage | The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% | | |
| | within the expiration date under appropriate storage condition. | | |
| | The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days, | | |
| | and compare O.D.values of the kit kept at 37C with that of at recommended temperature. (referring from China | | |
| | Biological Products Standard, which was calculated by the Arrhenius equation. For ELISA kit, 4 days storage | | |
| | at 37C can be considered as 6 months at 2 - 8C, which means 7 days at 37C equaling 12 months at 2 - 8C). | | |

| Application Details | | |
|------------------------------|-------------------------|--|
| Detect Range:0.156-10 ng/mL | | |
| Sensitivity:0.039 ng/mL | | |
| Sample Type:Serum, Plasma, G | other biological fluids | |
| Sample Volume: 1-200 µL | | |
| Assay Time:1-4.5h | | |
| Detection wavelength:450 nm | | |

Product Description

Detection Method:SandwichTest principle:This assay employs a two-site sandwich ELISA to quantitate 3-NT in samples. An antibody specific for 3-NT has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any3-NT present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for 3-NT is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of 3-NT bound in the initial step. The color development is stopped and the intensity of the color is measured.

Note: This product is for in vitro research use only