## Human Pregnancy Specific Protein B (PSPB) ELISA Kit

SAB Signalway Antibody

Catalog No: #EK8054

Package Size: #EK8054-1 48T #EK8054-2 96T

Orders: order@signalwayantibody.com Support: tech@signalwayantibody.com

## Description

Product Name	Human Pregnancy Specific Protein B (PSPB) ELISA Kit
Brief Description	ELISA Kit
Applications	ELISA
Species Reactivity	Human (Homo sapiens)
Other Names	FL-NCA-3; PSG; OTTHUMP00000200178 OTTHUMP00000200179 Pregnancy-specific
	beta-1-glycoprotein-5 pregnancy-specific beta 1 glycoprotein pregnancy-specific beta-1 glycoprotein
Accession No.	Q15238
Uniprot	Q15238
GeneID	5673;
Storage	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5%
	within the expiration date under appropriate storage condition.
	The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days,
	and compare O.D.values of the kit kept at 37C with that of at recommended temperature. (referring from China
	Biological Products Standard, which was calculated by the Arrhenius equation. For ELISA kit, 4 days storage
	at 37C can be considered as 6 months at 2 - 8C, which means 7 days at 37C equaling 12 months at 2 - 8C).

## **Application Details**

Detect Range:3.12-200 ng/mL	
Sensitivity:0.78 ng/mL	
Sample Type:Serum, Plasma, Other biological fluids	
Sample Volume: 1-200 μL	
Assay Time:1-4.5h	
Detection wavelength:450 nm	

## Product Description

Detection Method:SandwichTest principle:This assay employs a two-site sandwich ELISA to quantitate PSG5 in samples. An antibody specific for PSG5 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyPSG5 present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for PSG5 is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of PSG5 bound in the initial step. The color development is stopped and the intensity of the color is measured.

Note: This product is for in vitro research use only