## **Product Datasheet**

## Human Ninein (NIN) ELISA Kit

Catalog No: #EK8746

Package Size: #EK8746-1 48T #EK8746-2 96T



Orders: order@signalwayantibody.com Support: tech@signalwayantibody.com

Description	
Product Name	Human Ninein (NIN) ELISA Kit
Brief Description	ELISA Kit
Applications	ELISA
Species Reactivity	Human (Homo sapiens)
Other Names	KIAA1565; GSK3B-interacting protein glycogen synthase kinase 3 beta-interacting protein ninein ninein
	centrosomal protein
Accession No.	Q8N4C6
Uniprot	Q8N4C6
GeneID	51199;
Storage	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5%
	within the expiration date under appropriate storage condition.
	The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days,
	and compare O.D.values of the kit kept at 37C with that of at recommended temperature. (referring from China
	Biological Products Standard, which was calculated by the Arrhenius equation. For ELISA kit, 4 days storage
	at 37C can be considered as 6 months at 2 - 8C, which means 7 days at 37C equaling 12 months at 2 - 8C).

## **Application Details**

Detect Range:0.156-10 ng/mL		
Sensitivity:0.078 ng/mL		
Sample Type:Serum, Plasma, Other biological fluids		
Sample Volume: 1-200 μL		
Assay Time:1-4.5h		
Detection wavelength:450 nm		

## **Product Description**

Detection Method:SandwichTest principle:This assay employs a two-site sandwich ELISA to quantitate NIN in samples. An antibody specific for NIN has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyNIN present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for NIN is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of NIN bound in the initial step. The color development is stopped and the intensity of the color is measured.

Note: This product is for in vitro research use only