Mouse Nuclear receptor coactivator 2 (NCOA2) ELISA Kit

SAB Signalway Antibody

Catalog No: #EK8773

Package Size: #EK8773-1 48T #EK8773-2 96T

Orders: order@signalwayantibody.com Support: tech@signalwayantibody.com

Description

Product Name	Mouse Nuclear receptor coactivator 2 (NCOA2) ELISA Kit
Brief Description	ELISA Kit
Applications	ELISA
Species Reactivity	Mouse (Mus musculus)
Other Names	GRIP1; KAT13C; MGC138808; NCoA-2; SRC2; TIF2; bHLHe75; glucocorticoid receptor-interacting protein-1
Accession No.	Q61026
Uniprot	Q61026
GeneID	17978;
GeneID Storage	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5%
	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5%
	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition.
	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days,

Application Details

Detect Range:Request Information
sensitivity:Request Information
sample Type:Serum, Plasma, Other biological fluids
ample Volume: 1-200 μL
ssay Time:1-4.5h
Detection wavelength:450 nm

Product Description

Detection Method:SandwichTest principle:This assay employs a two-site sandwich ELISA to quantitate NCOA2 in samples. An antibody specific for NCOA2 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyNCOA2 present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for NCOA2 is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of NCOA2 bound in the initial step. The color development is stopped and the intensity of the color is measured.

Note: This product is for in vitro research use only