## Human N-acetyl-D-glucosamine kinase (NAGK) ELISA Kit

SAB Signalway Antibody

Catalog No: #EK8880

Package Size: #EK8880-1 48T #EK8880-2 96T

Orders: order@signalwayantibody.com Support: tech@signalwayantibody.com

## Description

Product Name	Human N-acetyl-D-glucosamine kinase (NAGK) ELISA Kit
Brief Description	ELISA Kit
Applications	ELISA
Species Reactivity	Human (Homo sapiens)
Other Names	GNK; HSA242910; N-Acetylglucosamine kinase
Accession No.	Q9UJ70
Uniprot	Q9UJ70
GeneID	55577;
Storage	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5%
	within the expiration date under appropriate storage condition.
	within the expiration date under appropriate storage condition.  The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days,
	· · · · · · · · · · · · · · · · · · ·
	The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days,

## **Application Details**

Detect Range:Request Information	
Sensitivity:Request Information	
Sample Type:Serum, Plasma, Other biological fluids	
Sample Volume: 1-200 μL	
Assay Time:1-4.5h	
Detection wavelength:450 nm	

## Product Description

Detection Method:SandwichTest principle:This assay employs a two-site sandwich ELISA to quantitate NAGK in samples. An antibody specific for NAGK has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyNAGK present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for NAGK is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of NAGK bound in the initial step. The color development is stopped and the intensity of the color is measured.

Note: This product is for in vitro research use only