## Human NEDD8-activating enzyme E1 regulatory subunit (NAE1) ELISA Kit

Signalway Antibody

Catalog No: #EK8895

Description

Package Size: #EK8895-1 48T #EK8895-2 96T

Orders: order@signalwayantibody.com Support: tech@signalwayantibody.com

Description	
Product Name	Human NEDD8-activating enzyme E1 regulatory subunit (NAE1) ELISA Kit
Brief Description	ELISA Kit
Applications	ELISA
Species Reactivity	Human (Homo sapiens)
Other Names	A-116A10.1; APPBP1; HPP1; ula-1; NEDD8-activating enzyme E1 subunit amyloid beta precursor protein
	binding protein 1; 59kDa amyloid beta precursor protein-binding protein 1; 59kD amyloid protein-bin
Accession No.	Q13564
Uniprot	Q13564
GeneID	8883;
Storage	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5%
	within the expiration date under appropriate storage condition.
	The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days,
	and compare O.D.values of the kit kept at 37C with that of at recommended temperature. (referring from China

## **Application Details**

Detect Range:Request Information
Sensitivity:Request Information
Sample Type:Serum, Plasma, Other biological fluids
Sample Volume: 1-200 μL
Assay Time:1-4.5h
Detection wavelength:450 nm

## Product Description

Detection Method:SandwichTest principle:This assay employs a two-site sandwich ELISA to quantitate NAE1 in samples. An antibody specific for NAE1 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyNAE1 present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for NAE1 is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of NAE1 bound in the initial step. The color development is stopped and the intensity of the color is measured.

Note: This product is for in vitro research use only

Biological Products Standard, which was calculated by the Arrhenius equation. For ELISA kit, 4 days storage at 37C can be considered as 6 months at 2 - 8C, which means 7 days at 37C equaling 12 months at 2 - 8C).