Rat Myosin-Ic (MYO1C) ELISA Kit

Catalog No: #EK8992

Signalway Antibo

Signalway Antibody

Orders: order@signalwayantibody.com Support: tech@signalwayantibody.com

Package Size: #EK8992-1 48T #EK8992-2 96T

De	sci	ip	tio	r

Product Name	Rat Myosin-Ic (MYO1C) ELISA Kit		
Brief Description	ELISA Kit		
Applications	ELISA		
Species Reactivity	Rat (Rattus norvegicus)		
Other Names	FLJ23903; MMI-beta; MMIb; NMI; myr2; myosin-l beta nuclear myosin l		
Accession No.	Q63355		
Uniprot	Q63355		
GeneID	65261;		
Storage	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5%		
	within the expiration date under appropriate storage condition.		
	The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days,		
	and compare O.D.values of the kit kept at 37C with that of at recommended temperature. (referring from China		
	Biological Products Standard, which was calculated by the Arrhenius equation. For ELISA kit, 4 days storage		
	at 37C can be considered as 6 months at 2 - 8C, which means 7 days at 37C equaling 12 months at 2 - 8C).		

Application Details

Detect Range:Request Information	
Sensitivity:Request Information	
Sample Type:Serum, Plasma, Other biological fluids	
Sample Volume: 1-200 μL	
Assay Time:1-4.5h	
Detection wavelength:450 nm	

Product Description

Detection Method:SandwichTest principle:This assay employs a two-site sandwich ELISA to quantitate MYO1C in samples. An antibody specific for MYO1C has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyMYO1C present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for MYO1C is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of MYO1C bound in the initial step. The color development is stopped and the intensity of the color is measured.

Note: This product is for in vitro research use only