Sheep Metallothionein-3 (MT3) ELISA Kit

Catalog No: #EK9269

Package Size: #EK9269-1 48T #EK9269-2 96T



Orders: order@signalwayantibody.com Support: tech@signalwayantibody.com

_		4.5	
Des	cri	ntid	าท
レしる	UH	ρ u	ווע

Sheep Metallothionein-3 (MT3) ELISA Kit	
ELISA Kit	
ELISA	
Sheep (Ovis aries)	
GIF; GIFB; GRIF; metallothionein 3 (growth inhibitory factor (neurotrophic))	
Q8MKE4	
Q8MKE4	
443168;	
The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5%	
within the expiration date under appropriate storage condition.	
The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days,	
and compare O.D.values of the kit kept at 37C with that of at recommended temperature. (referring from China	
Biological Products Standard, which was calculated by the Arrhenius equation. For ELISA kit, 4 days storage	
at 37C can be considered as 6 months at 2 - 8C, which means 7 days at 37C equaling 12 months at 2 - 8C).	

Application Details

Detect Range:Request Information	
Sensitivity:Request Information	
Sample Type:Serum, Plasma, Other biological fluids	
Sample Volume: 1-200 μL	
Assay Time:1-4.5h	
Detection wavelength:450 nm	

Product Description

Detection Method:SandwichTest principle:This assay employs a two-site sandwich ELISA to quantitate MT3 in samples. An antibody specific for MT3 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyMT3 present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for MT3 is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of MT3 bound in the initial step. The color development is stopped and the intensity of the color is measured.

Note: This product is for in vitro research use only