## Mouse Metallothionein-2 (MT-2) ELISA Kit

Catalog No: #EK9278

SAB Signalway Antibody

Package Size: #EK9278-1 48T #EK9278-2 96T

Orders: order@signalwayantibody.com Support: tech@signalwayantibody.com

| Description        |  |
|--------------------|--|
| Product Name       | Mouse Metallothionein-2 (MT-2) ELISA Kit   |
| Brief Description  | ELISA Kit  |
| Applications       | ELISA  |
| Species Reactivity | Mouse (Mus musculus)   |
| Storage            | The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% |
|                    | within the expiration date under appropriate storage condition.  |
|                    | The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days,      |
|                    | and compare O.D.values of the kit kept at 37C with that of at recommended temperature. (referring from China     |
|                    | Biological Products Standard, which was calculated by the Arrhenius equation. For ELISA kit, 4 days storage      |
|                    | at 37C can be considered as 6 months at 2 - 8C, which means 7 days at 37C equaling 12 months at 2 - 8C).         |

## **Application Details**

| Detect Range:Request Information                   |  |
|--|--|
| Sensitivity:Request Information                    |  |
| Sample Type:Serum, Plasma, Other biological fluids |  |
| Sample Volume: 1-200 μL                            |  |
| Assay Time:1-4.5h                                  |  |
| Detection wavelength:450 nm                        |  |

## **Product Description**

Detection Method:SandwichTest principle:This assay employs a two-site sandwich ELISA to quantitate MT-2 in samples. An antibody specific for MT-2 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyMT-2 present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for MT-2 is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of MT-2 bound in the initial step. The color development is stopped and the intensity of the color is measured.

Note: This product is for in vitro research use only