

# Human Methylthioribose-1-phosphate isomerase (MRI1) ELISA Kit



Catalog No: #EK9359

Orders: [order@signalwayantibody.com](mailto:order@signalwayantibody.com)

Package Size: #EK9359-1 48T #EK9359-2 96T

Support: [tech@signalwayantibody.com](mailto:tech@signalwayantibody.com)

## Description

Product Name	Human Methylthioribose-1-phosphate isomerase (MRI1) ELISA Kit
Brief Description	ELISA Kit
Applications	ELISA
Species Reactivity	Human (Homo sapiens)
Other Names	MGC3207; MTNA; Ypr118w; translation initiation factor eIF-2B subunit alpha/beta/delta-like protein
Accession No.	Q2NL31
Uniprot	Q2NL31
GeneID	534734;
Storage	<p>The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition.</p> <p>The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days, and compare O.D.values of the kit kept at 37C with that of at recommended temperature. (referring from China Biological Products Standard, which was calculated by the Arrhenius equation. For ELISA kit, 4 days storage at 37C can be considered as 6 months at 2 - 8C, which means 7 days at 37C equaling 12 months at 2 - 8C).</p>

## Application Details

Detect Range:Request Information

Sensitivity:Request Information

Sample Type:Serum, Plasma, Other biological fluids

Sample Volume: 1-200 µL

Assay Time:1-4.5h

Detection wavelength:450 nm

## Product Description

Detection Method:SandwichTest principle:This assay employs a two-site sandwich ELISA to quantitate MRI1 in samples. An antibody specific for MRI1 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyMRI1 present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for MRI1 is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of MRI1 bound in the initial step. The color development is stopped and the intensity of the color is measured.

Note: This product is for in vitro research use only