Mouse Menin (MEN1) ELISA Kit

Catalog No: #EK9705

Package Size: #EK9705-1 48T #EK9705-2 96T



Orders: order@signalwayantibody.com Support: tech@signalwayantibody.com

_			
	escri	Intic	۱n
			и і

Product Name	Mouse Menin (MEN1) ELISA Kit	
Brief Description	ELISA Kit	
Applications	ELISA	
Species Reactivity	Mouse (Mus musculus)	
Other Names	MEAI; SCG2; OTTHUMP00000070176 OTTHUMP00000070184 endocrine adenomatosis; multiple menin	
Accession No.	O88559	
Uniprot	O88559	
GeneID	17283;	
Storage	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5%	
	within the expiration date under appropriate storage condition.	
	The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days,	
	and compare O.D.values of the kit kept at 37C with that of at recommended temperature. (referring from China	
	Biological Products Standard, which was calculated by the Arrhenius equation. For ELISA kit, 4 days storage	
	at 37C can be considered as 6 months at 2 - 8C, which means 7 days at 37C equaling 12 months at 2 - 8C).	

Application Details

Detect Range:Request Information		
Sensitivity:Request Information		
Sample Type:Serum, Plasma, Other biological fluids		
Sample Volume: 1-200 μL		
Assay Time:1-4.5h		
Detection wavelength:450 nm		

Product Description

Detection Method:SandwichTest principle:This assay employs a two-site sandwich ELISA to quantitate MEN1 in samples. An antibody specific for MEN1 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyMEN1 present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for MEN1 is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of MEN1 bound in the initial step. The color development is stopped and the intensity of the color is measured.

Note: This product is for in vitro research use only