## **Product Datasheet**

## Human Mediator of RNA polymerase II transcription subunit 12 (MED12) ELISA Kit



Catalog No: #EK9743

Description

Package Size: #EK9743-1 48T #EK9743-2 96T

Orders: order@signalwayantibody.com Support: tech@signalwayantibody.com

Product Name	Human Mediator of RNA polymerase II transcription subunit 12 (MED12) ELISA Kit
Brief Description	ELISA Kit
Applications	ELISA
Species Reactivity	Human (Homo sapiens)
Other Names	ARC240; CAGH45; FGS1; HOPA; KIAA0192; OKS; OPA1; TNRC11; TRAP230; CAG repeat protein
	45 OPA-containing protein activator-recruited cofactor 240 kDa component human opposite paired mediator of
	RNA po
Accession No.	Q93074
Uniprot	Q93074
GeneID	9968;
Storage	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5%

within the expiration date under appropriate storage condition.

The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days, and compare O.D.values of the kit kept at 37C with that of at recommended temperature. (referring from China Biological Products Standard, which was calculated by the Arrhenius equation. For ELISA kit, 4 days storage at 37C can be considered as 6 months at 2 - 8C, which means 7 days at 37C equaling 12 months at 2 - 8C).

## **Application Details**

Detect Range:Request Information
Sensitivity:Request Information
Sample Type:Serum, Plasma, Other biological fluids
Sample Volume: 1-200 μL
Assay Time:1-4.5h
Detection wavelength:450 nm

## **Product Description**

Detection Method:SandwichTest principle:This assay employs a two-site sandwich ELISA to quantitate MED12 in samples. An antibody specific for MED12 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyMED12 present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for MED12 is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of MED12 bound in the initial step. The color development is stopped and the intensity of the color is measured.

Note: This product is for in vitro research use only