Product Datasheet

Human Trans-1,2-dihydrobenzene-1,2-diol dehydrogenase (DHDH) ELISA Kit

Catalog No: #EK10740

Package Size: #EK10740-1 48T #EK10740-2 96T



Orders: order@signalwayantibody.com Support: tech@signalwayantibody.com

D	escri	pt	ioi	1

Product Name	Human Trans-1,2-dihydrobenzene-1,2-diol dehydrogenase (DHDH) ELISA Kit
Brief Description	ELISA Kit
Applications	ELISA
Species Reactivity	Human (Homo sapiens)
Other Names	2DD; HUM2DD; 3-deoxyglucosone reductase D-xylose 1-dehydrogenase D-xylose-NADP
	dehydrogenase dimeric dihydrodiol dehydrogenase
Accession No.	Q9UQ10
Uniprot	Q9UQ10
GeneID	27294;
Storage	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5%
	within the expiration date under appropriate storage condition.
	The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days,
	and compare O.D.values of the kit kept at 37C with that of at recommended temperature. (referring from China
	Biological Products Standard, which was calculated by the Arrhenius equation. For ELISA kit, 4 days storage
	at 37C can be considered as 6 months at 2 - 8C, which means 7 days at 37C equaling 12 months at 2 - 8C).

Application Details

Detect Range:Request Information		
Sensitivity:Request Information		
Sample Type:Serum, Plasma, Other biological fluids		
Sample Volume: 1-200 μL		
Assay Time:1-4.5h		
Detection wavelength:450 nm		

Product Description

Detection Method:SandwichTest principle:This assay employs a two-site sandwich ELISA to quantitate DHDH in samples. An antibody specific for DHDH has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyDHDH present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for DHDH is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of DHDH bound in the initial step. The color development is stopped and the intensity of the color is measured.

Note: This product is for in vitro research use only