## Human DEP domain-containing mTOR-interacting protein (DEPDC6) ELISA Kit

Signalway Antibody

Catalog No: #EK10795

Package Size: #EK10795-1 48T #EK10795-2 96T

Orders: order@signalwayantibody.com Support: tech@signalwayantibody.com

D	esci	rıpt	ion

Product Name	Human DEP domain-containing mTOR-interacting protein (DEPDC6) ELISA Kit	
Brief Description	ELISA Kit	
Applications	ELISA	
Species Reactivity	Human (Homo sapiens)	
Other Names	DEP.6; DEPTOR; DKFZp564B1778; FLJ12341; FLJ12428; FLJ13854; DEP domain containing mTOR	
	interacting protein	
Accession No.	Q8TB45	
Uniprot	Q8TB45	
GeneID	64798;	
Storage	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5%	
	within the expiration date under appropriate storage condition.	
	The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days,	
	and compare O.D.values of the kit kept at 37C with that of at recommended temperature. (referring from China	
	Biological Products Standard, which was calculated by the Arrhenius equation. For ELISA kit, 4 days storage	
	at 37C can be considered as 6 months at 2 - 8C, which means 7 days at 37C equaling 12 months at 2 - 8C).	

## **Application Details**

Detect Range:Request Information
Sensitivity:Request Information
Sample Type:Serum, Plasma, Other biological fluids
Sample Volume: 1-200 μL
Assay Time:1-4.5h
Detection wavelength:450 nm

## **Product Description**

Detection Method:SandwichTest principle:This assay employs a two-site sandwich ELISA to quantitate DEPDC6 in samples. An antibody specific for DEPDC6 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyDEPDC6 present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for DEPDC6 is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of DEPDC6 bound in the initial step. The color development is stopped and the intensity of the color is measured.

Note: This product is for in vitro research use only