## Human Cytokeratin fragment antigen 21-1 (CYFRA21-1) ELISA Kit

Catalog No: #EK10987

Description

Package Size: #EK10987-1 48T #EK10987-2 96T



Orders: order@signalwayantibody.com Support: tech@signalwayantibody.com

Beeenparen				
Product Name	Human Cytokeratin fragment antigen 21-1 (CYFRA21-1) ELISA Kit			
Brief Description	ELISA Kit			
Applications	ELISA			
Species Reactivity	Human (Homo sapiens)			
Storage	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5%			
	within the expiration date under appropriate storage condition.			
	The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days,			
	and compare O.D.values of the kit kept at 37C with that of at recommended temperature. (referring from China			
	Biological Products Standard, which was calculated by the Arrhenius equation. For ELISA kit, 4 days storage			
	at 37C can be considered as 6 months at 2 - 8C, which means 7 days at 37C equaling 12 months at 2 - 8C).			

Application Details				
Detect Range:Request Information				
Sensitivity:Request Information				
Sample Type:Serum, Plasma, Other biological fluids				
Sample Volume: 1-200 µL				
Assay Time:1-4.5h				
Detection wavelength:450 nm				

## **Product Description**

Detection Method:SandwichTest principle:This assay employs a two-site sandwich ELISA to quantitate CYFRA21-1 in samples. An antibody specific for CYFRA21-1 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyCYFRA21-1 present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for CYFRA21-1 is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of CYFRA21-1 bound in the initial step. The color development is stopped and the intensity of the color is measured.

Note: This product is for in vitro research use only