Human Dengue fever antibody (DF-Ab) ELISA Kit

Catalog No: #EK11739

Package Size: #EK11739-1 48T #EK11739-2 96T



Orders: order@signalwayantibody.com Support: tech@signalwayantibody.com

Description

| Product Name | Human Dengue fever antibody (DF-Ab) ELISA Kit | | |
|--------------------|--|--|--|
| Brief Description | ELISA Kit | | |
| Applications | ELISA | | |
| Species Reactivity | Human (Homo sapiens) | | |
| Storage | The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% | | |
| | within the expiration date under appropriate storage condition. | | |
| | The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days, | | |
| | and compare O.D.values of the kit kept at 37C with that of at recommended temperature. (referring from China | | |
| | Biological Products Standard, which was calculated by the Arrhenius equation. For ELISA kit, 4 days storage | | |
| | at 37C can be considered as 6 months at 2 - 8C, which means 7 days at 37C equaling 12 months at 2 - 8C). | | |
| | | | |

| Application Details | | | | |
|--|--|--|--|--|
| Detect Range:Detection antibody | | | | |
| Sensitivity:Request Information | | | | |
| Sample Type:Serum, Plasma, Other biological fluids | | | | |
| Sample Volume: 1-200 µL | | | | |
| Assay Time:1-4.5h | | | | |
| Detection wavelength:450 nm | | | | |

Product Description

Detection Method:Competitive ELISATest principle:This assay employs the competitive enzyme immunoassay technique. The microtiter plate provided in this kit has been pre-coated with an antibody specific to DF-Ab. Standards or samples are then added to the appropriate microtiter plate wells with a Horseradish Peroxidase (HRP)-conjugated DF-Ab and incubated. The competitive inhibition reaction is launched between with HRP labeled DF-Ab and unlabeled DF-Ab with the antibody. A substrate solution is added to the wells and the color develops in opposite to the amount of DF-Ab in the sample. The color development is stopped and the intensity of the color is measured.

Note: This product is for in vitro research use only