Product Datasheet

Human Dentin Matrix Protein 1 (DMP1) ELISA Kit

Catalog No: #EK12138



Package Size: #EK12138-1 48T #EK12138-2 96T

Orders: order@signalwayantibody.com Support: tech@signalwayantibody.com

| Description | | |
|--------------------|--|--|
| Product Name | Human Dentin Matrix Protein 1 (DMP1) ELISA Kit | |
| Brief Description | ELISA Kit | |
| Applications | ELISA | |
| Species Reactivity | Human (Homo sapiens) | |
| Other Names | ARHP; ARHR; DMP-1; dentin matrix protein 1 | |
| Accession No. | Q13316 | |
| Uniprot | Q13316 | |
| GeneID | 1758; | |
| Storage | The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% | |
| | within the expiration date under appropriate storage condition. | |
| | The loss rate was determined by accelerated thermal degradation test. Keep the kit at 37C for 4 and 7 days, | |
| | and compare O.D.values of the kit kept at 37C with that of at recommended temperature. (referring from China | |
| | Biological Products Standard, which was calculated by the Arrhenius equation. For ELISA kit, 4 days storage | |
| | at 37C can be considered as 6 months at 2 - 8C, which means 7 days at 37C equaling 12 months at 2 - 8C). | |

| App | dication | Details |
|-----|----------|---------|
| | | |

| Detect Range:0.31-20 ng/mL | |
|--|--|
| Sensitivity:0.156 ng/mL | |
| Sample Type:Serum, Plasma, Other biological fluids | |
| Sample Volume: 1-200 μL | |
| Assay Time:1-4.5h | |
| Detection wavelength:450 nm | |

Product Description

Detection Method:SandwichTest principle:This assay employs a two-site sandwich ELISA to quantitate DMP1 in samples. An antibody specific for DMP1 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and anyDMP1 present is bound by the immobilized antibody. After removing any unbound substances, a biotin-conjugated antibody specific for DMP1 is added to the wells. After washing, Streptavidin conjugated Horseradish Peroxidase (HRP) is added to the wells. Following a wash to remove any unbound avidin-enzyme reagent, a substrate solution is added to the wells and color develops in proportion to the amount of DMP1 bound in the initial step. The color development is stopped and the intensity of the color is measured.Product Overview:DMP1?is critical for proper mineralization of bone and dentin, is present in diverse cells of bone and tooth tissues. The protein contains a large number of acidic domains, multiple phosphorylation sites, a functional arg-gly-asp cell attachment sequence, and a DNA binding domain. In undifferentiated osteoblasts it is primarily a nuclear protein that regulates the expression of osteoblast-specific genes. During osteoblast maturation the protein becomes phosphorylated and is exported to the extracellular matrix, where it orchestrates mineralized matrix formation. Mutations in the gene are known to cause autosomal recessive hypophosphatemia, a disease that manifests as rickets and osteomalacia. The gene structure is conserved in mammals. Two transcript variants encoding different isoforms have been described for this gene.

Note: This product is for in vitro research use only